






# Aries Medical Management

Dongtai Huayi / Perfect Fit

Disposable Nitrile Examination Glove | 510K + Chemo Rated



DONG TAI CITY HUAYI GLOVES CO.,LTD  
NO.36 NANZHANG ROAD,DONGTAI HIGH-TECH  
INDUSTRIAL DEVELOPMENT ZONE JIANGSU PROVINCE

The following sample(s) was/were submitted and identified by the client as:

Sample Description : DISPOSABLE NITRILE GLOVES  
Lot No. : NOT PROVIDED  
Lot Size : NOT PROVIDED  
Sample Quantity : 300PCS  
Sample Receiving Date : JUN.28,2020  
Final Information Date : JUN.30,2020  
Testing Period : JUN.28,2020 TO JUL.15,2020

<u>Test Requested</u>	<u>Result</u>
ASTM D6319-2019 Standard Specification for Nitrile Examination Gloves for Medical Application (Clause 5, 6.1.2, 6.1.3, 6.1.4, 6.1.5)	Pass

Remark: Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.

SGS-CSTC Standards  
Technical Services (Qingdao)  
Co., Ltd.

*Jessica Gao*



\_\_\_\_\_  
Jessica Gao  
Approved Signatory



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/sgs-terms-and-conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/sgs-terms-and-conditions/terms-e-documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues contained therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not constitute parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.  
Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755)82071442, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

SGS-CSTC Standards Technical Services (Qingdao) Co., Ltd.

SGS Center, No. 143, Zhuzhou Road, Laoshan District, Qingdao, China 266101  
T: (86-532) 88999888 | (86-532) 80991955

[www.sgs.com](http://www.sgs.com)  
[sgs.china@sgs.com](mailto:sgs.china@sgs.com)

**Test Conducted:**
**ASTM D6319-2019 Standard Specification for Nitrile Examination Gloves for Medical Application**

Number of test sample	:	231 Pieces
Size	:	L

Clause	Test Items	Result
5	Sampling	See Result 1
6	Performance Requirements	---
6.1.2	Freedom from Holes	Pass (See Result 2)
6.1.3	Physical dimensions	Pass (See Result 3)
6.1.4	Physical property characteristics	Pass (See Result 4)
6.1.5	Powder Residue for Powder Free Gloves	Pass (See Result 5)

**Test Result:**
**Result 1: Sampling**

The number of specimen:

	Sample size	Ac	Re
Freedom from Holes	200	10	11
Dimensions	13	1	2
Physical property	13	1	2

**Result 2: Freedom from Holes**

Sample Quantity: 200 Pieces

AQL=2.5      Ac: 10      Re: 11      Found: 0



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/terms-conditions> and, for electronic format documents, subject to terms and Conditions for Electronic Documents at <http://www.sgs.com/en/terms-conditions/terms-document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from ensuring all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: [CN.Descheck@sgs.com](mailto:CN.Descheck@sgs.com)

SGS Center, No. 143, Zhuzhou Road, Laxshen District, Qingdao, China 266101  
 T (86-532) 58999885 F (86-532) 81991955

[www.sgs.com.cn](http://www.sgs.com.cn)  
[sgs.china@sgs.com](mailto:sgs.china@sgs.com)

**Result 4: Physical property characteristics**

Sample Quantity: 13 Pieces

AQL=4.0      Ac: 1      Re: 2

Size: L					
Before Aging			After Aging		
Sample No.	Tensile strength (Mpa)	Ultimate Elongation (%)	Sample No.	Tensile strength (Mpa)	Ultimate Elongation (%)
1	30.8	557	1	27.5	509
2	27.6	550	2	31.2	513
3	30.8	558	3	31.3	526
4	28.6	548	4	23.2	493
5	31.0	557	5	28.5	513
6	30.3	552	6	27.6	505
7	30.7	557	7	26.4	498
8	28.0	540	8	28.8	506
9	27.3	538	9	29.0	514
10	29.3	544	10	23.7	492
11	32.0	551	11	28.6	517
12	25.0	530	12	30.1	489
13	30.5	545	13	34.0	517
Standard requirement	≥14	≥500	Standard requirement	≥14	≥400
Found	0	0	Found	0	0

**Result 5: Powder Residue For Powder Free Gloves**

The average mass per glove(mg)	0.02
Standard requirement(mg)	≤2.0

Remark: The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.



**Result 3: Physical dimensions**

Sample Quantity: 13 Pieces

AQL=4.0

Ac:1

Re: 2

Sample No.	Size: L			
	Length/mm	Width/mm	Median value/mm	
			Thickness-finger	Thickness-palm
1	245	110	0.131	0.104
2	245	110	0.138	0.108
3	245	110	0.133	0.104
4	240	110	0.131	0.107
5	243	111	0.131	0.106
6	243	110	0.131	0.107
7	245	110	0.138	0.108
8	242	110	0.138	0.105
9	240	110	0.132	0.103
10	242	110	0.133	0.093
11	241	110	0.134	0.105
12	240	110	0.135	0.104
13	240	110	0.134	0.104
Standard requirement	≥230	110±10	≥0.05	≥0.05
Found	0	0	0	0



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at [http://www.sgs.com/eng/terms\\_and\\_conditions\\_sgs.asp](http://www.sgs.com/eng/terms_and_conditions_sgs.asp) and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at [http://www.sgs.com/eng/terms\\_and\\_conditions/terms\\_e-document.asp](http://www.sgs.com/eng/terms_and_conditions/terms_e-document.asp). Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing/inspection report & certificate, please contact us at telephone: (86-755)83071443, or email: [CN\\_Dispatch@sgs.com](mailto:CN_Dispatch@sgs.com)

SGS (China) Technical Services Co., Ltd.  
 18888888888

SGS Center, No. 143, Zhuzhou Road, Laxian District, Qingdao, China 266101  
 t (86-532) 88969888 f (86-532) 80991955

[www.sgs.com](http://www.sgs.com)  
[sgs.china@sgs.com](mailto:sgs.china@sgs.com)

Sample Photo:

Received sample



SGS authenticate the photo on original report only



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/terms-and-conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/terms-and-conditions/terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Attention: To check the authenticity of testing inspection report & certificate, please contact us at telephone: (86-755)83271443, or email: [CN.Questech@sgs.com](mailto:CN.Questech@sgs.com)

SGS Center, No. 143, Zhuzhou Road, Luohu District, Qingdao, China 266101  
T: (86-532) 68999888 F: (86-532) 60991955

[www.sgs.com.cn](http://www.sgs.com.cn)  
[sgs.china@sgs.com](mailto:sgs.china@sgs.com)





# EC REP CERTIFICATE



## CMC MEDICAL DEVICES & DRUGS SL NO. CMC/CE/2020/15062020.37

CONFIRMED THAT CMC MEDICAL DEVICES & DRUGS S.L. Is the European Authorized Representative of

**DongTai City HuaYi Gloves CO.,LTD.**  
NO. 36 Nanzhuang Road, High-Tech Industrial Development Zone of  
Dongtai City, Jiangsu Province, CHINA

The certificate remains valid until the expiration agreement of EC REP, manufacturing conditions, the quality system or relevant legislation are changed. The validity is conditioned by positive results of periodic surveillance audits.

The product liability rests with the manufacturer in accordance with applicable directive and standard, after fulfilling of the relevant EU legislation requirements, the manufacturer shall affix relevant CE marking to all above mentioned models of the medical device.

Complies with the applicable essential requirements of the council directive 93/42/EEC on medical devices as amended.

The products in Annex I was registered in Spanish MOH with number **RPS/1418/2020**



Issued on: 15/06/2020

  
Authorized Signatory  
CMC Medical Devices & Drugs SL

Valid until: 14/06/2021


# EC REP CERTIFICATE



ANNEX I Medical Device Products



Nitrile Examination Gloves



# CE

**Section 6 510(k) Summary****510(K) Summary**

"This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR 807.92."

"The assigned 510(k) number is: K131823 "

**Premarket Notification [510(k)] Summary****1.0 Submitter:**

Submitter's name : Dong Tai City Huayi Gloves Co., Ltd.  
 Submitter's address : No.68 Jingyi Road, ChengDong New District, DongTai, Jiangsu, 224200, China  
 Phone number : 0086-515-85332088  
 Fax number : 0086-515-85332688  
 Name of contact person: Ms.Zhu Ya Fen  
 Date of preparation : 2013-10-07

**2.0 Name of the Device**

Device Name: Powder Free Nitrile Patient Examination Gloves, Blue Color  
 Proprietary/Trade name: DongTai  
 Common Name: Exam gloves  
 Classification Name: Patient examination glove  
 Device Classification: I  
 Regulation Number: 21 CFR 880.6250  
 Panel: General Hospital (80)  
 Product Code: LZA

**3.0 Predicate device**

Device Name: Powder Free Nitrile Patient Examination Gloves, Blue Color  
 Company name: Jiangsu Dongling Plastic & Rubber Co., Ltd.  
 510(K) Number: K110247

**4.0 Device Description:**

**4.1 How the device functions:**  
 Nitrile films form a barrier to body fluids and bloodborne Pathogens

**4.2 Scientific concepts that form the basis for the device**

The nitrile rubber is water tight under normal conditions of use. Its tensile properties cause it to conform to the hand, allowing movements necessary for a medical procedure.

**4.3 Physical and performance characteristics such as design, materials and physical properties:**

Nitrile glove is known to create a barrier to bloodborne pathogens and body fluids. ASTM conforming tensile properties create a glove that is strong and flexible. The leaching process removes traces of chemical accelerants that may be chemically irritating. The glove is manufactured in accordance with the requirements of ASTM D6319 and ASTM D5151 requirements.

**5.0 Device Intended Use (Indication for use):**

Powder Free Nitrile Patient Examination Gloves, Blue Color is a non-sterile disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

**6.0 Summary of the Technological Characteristics of the Device:**

The Powder Free Nitrile Patient Examination Gloves, Blue Color, non sterile are summarized with the following technological characteristics compared to ASTM or equivalent standard.

Characteristics	Standard	Device performance
Dimension	ASTM standard D 6319-10.	Meets
Physical Properties	ASTM standard D 6319-10.	Meets
Freedom from pinholes	21 CFR 800.20	Meets
Powder Residual	ASTM standard D 6319-10 and D6124-06(Reapproved 2011).	Meets <2mg/glove
Biocompatibility	Primary Skin Irritation in rabbits ISO 10993-10:2002 /Amd.1:2006	Passes Not a Primary Skin Irritant
	Dermal sensitization in the guinea pig ISO 10993-10: 2002 /Amd.1:2006	Passes Not a Dermal Sensitizer

**7.0 Substantial Equivalent Based on Assessment of Non-Clinical Performance Data:**

Powder Free Nitrile Patient Examination Gloves, Blue Color, meet requirements per ASTM D6319-10, per ASTM D6124-06(Reapproved 2011), per 21 CFR 800.20 and ISO 10993-10: 2002/Amd.1:2006.

The performance test data of the non clinical tests that support a determination of substantial equivalent is the same as mentioned immediately above.

**8.0 Substantial Equivalent Based on Assessment of Clinical Performance Data:**

Clinical data is not needed for gloves or for most devices cleared by the 510(k) process.

**9.0 Substantial Equivalence Comparison:**

Features & Description	Predicate Device	Medical Glove Guidance Manual	Subject Device	Result of Comparison
Company	Jiangsu Dongling Plastic & Rubber Co., Ltd.		Dong Tai City Huayi Gloves Co., Ltd.	--
510(K) Number	K110247		K131823	
Product name	Powder Free Nitrile Patient Examination Gloves, Blue Color		Powder Free Nitrile Patient Examination Gloves, Blue Color	same
Product Code	LZA	LZA	LZA	same
Size	Small/ Medium/ Large/X large		Small/ Medium/ Large/X large	same
Intend for use	Powder Free Nitrile Patient Examination Gloves, Blue Color is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.	Powder Free Examination Gloves: A powder-free patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.	Powder Free Nitrile Patient Examination Gloves, Blue Color is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.	Substantially equivalent
Device Description and Specifications	Meets ASTM D6319-10	If nitrile gloves: Does the above data for your nitrile examination glove meet all the current specifications listed under the applicable ASTM standard D6319 or an equivalent consensus standard?	Meets ASTM D6319-10	Substantially equivalent
Dimensions --Length	Meets ASTM D6319-10 ≥230mm min	ASTM D6319	230mm min for all sizes	Substantially equivalent

Dimensions -- Width	Meets ASTM D6319-10  Small 70-90 mm Medium 85-105mm Large 100-120mm Xlarge 110-130 mm	ASTM D6319	Small 82-86 mm Medium 94-98mm Large 107-113mm X large 115-121 mm	Substantially equivalent
Dimensions --Thickness	Meets ASTM D6319-10  Finger 0.05mm min. Palm 0.05mm min.		Thickness (mm) min. Finger 0.09-0.13 Palm 0.09-0.13	Substantially equivalent
Physical Properties	Meets ASTM D D6319-10  Before aging/after aging Elongation ≥500% Tensile Strength≥ 14MPa	ASTM D6319	Before aging/after aging  Elongation :540-600% Tensile Strength:19-23 MPa	Substantially equivalent
Freedom from Pinholes	Meets • 21 CFR 800.20 • ASTM D6319-10 • ASTM D 5151-06 (Reapproved 2011)	21 CFR 800.20 ASTM D5250 ASTM D 5151	Meets ASTM D5151-06 (Reapproved 2011)  Holes at Inspection Level 1 AQL2.5	Substantially equivalent
Residual Powder	Meets ASTM D 6124-06 (Reapproved 2011)  below 2mg of residual powder	ASTM D 6124	Meets ASTM D 6124-06 (Reapproved 2011)  Results generated values below 2mg of residual powder	Substantially equivalent
Materials used to fabricate the devices	Nitrile	If the glove is made of a polymer or other type of material, identify the material.	Nitrile	Substantially equivalent
Dusting or Donning Powder:	PU	If a donning lubricant is used, state the composition and include biocompatibility data for the lubricant in an identified attachment; also state the name, manufacturer, and address below	PU-120C	Substantially equivalent
Dusting or Donning Powder: name	PU	Lubricant Generic Name/ Lubricant Brand Name	Surface Coating Agent	Substantially equivalent
Compare performance data supporting substantial equivalence	Meets ASTM D5151-06 (Reapproved 2011) ASTM D6319-10 ASTM D6124-06 (Reapproved 2011)	At this time FDA recognizes the following standards: Patient Examination Gloves(PVC)ASTM D5151(Detection of Holes in Medical Gloves)ASTM D6124(Residual Powder on Medical Gloves)ASTM	Meets ASTM D5151-06 (Reapproved 2011) ASTM D6319-10 ASTM D6124-06 (Reapproved 2011)	Substantially equivalent

		D6319 (Nitrile Gloves)		
Single Patient Use	Single Patient Use	Single Patient Use	Single Patient Use	Substantially equivalent
Biocompatibility	SKIN IRRITATION DERMAL and SENSITIZATION STUDIES Meets ISO 10993-10:2002/Amd.1:2006	SKIN IRRITATION DERMAL and SENSITIZATION STUDIES ISO 10993-10	The test article was a non-irritant and non-sensitizer.  SKIN IRRITATION DERMAL and SENSITIZATION STUDIES Meets ISO 10993-10:2002 /Amd.1:2006	Substantially equivalent
Labeling for the legally marketed device to which substantial equivalence is claimed.	-Powder Free -Patient Examination Glove -Single Use Only - Manufactured For: - Lot - Blue color - Non-sterile	Chapter 4	-Powder Free -Patient Examination Glove -Single Use Only - Manufactured For: - Lot - Blue color - Non-sterile	Substantially equivalent

**10.0 Substantial Equivalence Comparison:**

It can be concluded that the Powder Free Nitrile Patient Examination Gloves, Blue Color meet the ASTM standard or equivalent standard and FDA requirements for waterleak test on pinhole AQL., meet labeling claims.

It can be concluded that the Powder Free Nitrile Patient Examination Gloves, Blue Color is as safe, as effective, and performs as well as the predicate device, Powder Free Nitrile Patient Examination Gloves, Blue Color, Jiangsu Dongling Plastic & Rubber Co., Ltd. K110247



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center - WO66-G609  
Silver Spring, MD 20993-0002

December 9, 2013

Dong Tai City Huayi Gloves Company, Limited  
C/O Mr. Chu Xiaoran  
Official Correspondent  
Beijing Easy-Link Co.  
Room 1606 Bldg. 1 Jianxiang Yuan #209  
Bei Si Huan Zhong Road, Haidian District  
Beijing 100083  
CHINA

Re: K131823

Trade/Device Name: Powder-Free Nitrile Patient Examination Gloves, Blue Color  
Regulation Number: 21 CFR 880.6250  
Regulation Name: Patient Examination Glove  
Regulatory Class: I  
Product Code: LZA  
Dated: October 7, 2013  
Received: October 24, 2013

Dear Mr. Xiaoran:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.



Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Erin I. Keith, M.D.  
Clinical Assistant Director

FOR

Erin I. Keith, M.S.  
Acting Director  
Division of Anesthesiology, General Hospital,  
Respiratory, Infection Control and  
Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

**Indications for Use**

510(k) Number (if known)  
K131823

Device Name  
Powder Free Nitrile Patient Examination Gloves, Blue Color

Indications for Use (Describe)

Powder Free Nitrile Patient Examination Gloves, Blue Color is a non-sterile disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)  Over-The-Counter Use (21 CFR 801 Subpart C)

PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON A SEPARATE PAGE IF NEEDED.

FOR FDA USE ONLY

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

Elizabeth F. Claverie -S  
2013.12.06 15:33:05 -05'00'

June 9,2021

• **TEST REPORT** •

**PN 13326**


**CHEMICAL ANALYTICAL SERVICES**

**DongTai City HuaYi Gloves CO.,Ltd  
NO.36 Nanzhuang Road,High-Tech Industrial  
Development Zone of Dongtai City  
Jiangsu Provice,224249 P.R.  
China**

Prepared By:

  
Tiffany L. Heller  
Assistant Manager, Pharmaceutical Services

Approved By:

  
Ana C. Barbur, M.S.  
Manager, Chemical, Microbiological, & Pharmaceutical Services



An A2LA ISO 17025 Accredited Testing Laboratory — Certificate Numbers 255.01 & 255  
ISO 9001:2008 Registered

**ISO 9001:2008**  
Registered

Letters and reports are for the exclusive use of the clients to whom they are addressed and shall not be reproduced, except in full, without the written permission of Akron Rubber Development Laboratory, Inc. (ARDL). The information contained herein applies to the specific material, products or processes tested or evaluated. No warranty of any kind is herein construed or implied. The liability of ARDL, Inc. shall be limited to the amount of consideration paid for services. ARDL, Inc. is ISO 17025 accredited by A2LA for the test methods listed on the referenced certificates.

June 9,2021

**DongTai City HuaYi Gloves CO.,Ltd.**

Page 1 of 3 – PN 13326

**SUBJECT:** Permeation testing per ASTM D 6978-05 on sample submitted by the above company.

**RECEIVED:** Glove sample identified as Nitrile Powder Free Glove, Blue, Lot# ZP161212.

**TESTING CHEMOTHERAPY DRUGS:**

Table 1. List of the Testing Chemotherapy Drugs, Sources, and Expiration Dates

TESTING CHEMOTHERAPY DRUGS	DRUG SOURCE
Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)	Sigma Aldrich; Lot# 016M4028V; Expires 09/2017
Cisplatin, 1.0 mg/ml (1,000 ppm)	Fresenius Kabi; Lot# 6114286; 01/2018
Cyclophosphamide (Cytosan), 20.0 mg/ml (20,000ppm)	Sigma Aldrich; Lot# BCBM8984V; Expiration 04/2017
Cytarabine, 100 mg/ml (100,000 ppm)	Sigma Aldrich; Lot# LRAA8717; Expiration 01/2018
Dacarbazine (DTIC), 10.0 mg/ml (10,000 ppm)	Teva; Lot# 31318323B; Expiration 10/8/2017
Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm)	Sigma Aldrich; Lot# SLBM7382V; Expiration 08/2017
Etoposide (Toposar), 20.0 mg/ml (20,000 ppm)	Teva; Lot# 31321666B; Expiration 09/2019
Fluorouracil, 50.0 mg/ml (50,000 ppm)	Accord; Lot# PT04300; Expiration 10/2018
Ifosfamide, 50.0 mg/ml (50,000 ppm)	Sigma Aldrich; Lot# 106K1063V; Expiration 12/2017
Methotrexate, 25 mg/ml, (25,000 ppm)	Teva; Lot# 16A28MA, Expiration 01/2018
Mitomycin C, 0.5 mg/ml (500 ppm)	Sigma; Lot# MKBR2210V; Expiration 03/2017
Mitoxantrone, 2.0mg/ml (2,000ppm)	Sigma Aldrich; Lot# MKBR2210V; Expiration 11/2017
Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm)	Hospira; Lot# C126865AA; Expiration 12/2017
Thiotepa, 10.0 mg/ml (10,000 ppm)	Sigma Aldrich; Lot# SLBQ8871V; Expiration 02/2018
Vincristine Sulfate, 1.0 mg/ml (1,000 ppm)	Sigma Aldrich; Lot# SLBQ9329V; Expiration 01/2018

**COLLECTION MEDIA:**

The collection media, which were selected, are listed in Table 2.

Table 2. Collection Media for Testing Chemotherapy Drugs

TEST DRUG AND CONCENTRATION	COLLECTION MEDIUM
Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)	10% Ethanol Aqueous Solution
Cisplatin, 1.0 mg/ml (1,000 ppm)	Distilled Water
Cyclophosphamide (Cytosan), 20.0 mg/ml (20,000ppm)	Distilled Water
Cytarabine, 100 mg/ml (100,000 ppm)	Distilled Water
Dacarbazine (DTIC), 10.0 mg/ml (10,000 ppm)	Distilled Water
Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm)	Distilled Water
Etoposide (Toposar), 20.0 mg/ml (20,000 ppm)	Distilled Water
Fluorouracil, 50.0 mg/ml (50,000 ppm)	9.20 pH Sodium Hydroxide Solution
Ifosfamide, 50.0 mg/ml (50,000 ppm)	Distilled Water
Methotrexate, 25 mg/ml, (25,000 ppm)	Distilled Water
Mitomycin C, 0.5 mg/ml (500 ppm)	Distilled Water
Mitoxantrone, 2.0mg/ml (2,000ppm)	Distilled Water
Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm)	30% Methanol Aqueous Solution
Thiotepa, 10.0 mg/ml (10,000 ppm)	Distilled Water
Vincristine Sulfate, 1.0 mg/ml (1,000 ppm)	Distilled Water

**DongTai City HuaYi Gloves CO.,Ltd.**

Page 2 of 3 – PN 13326

**TESTING CONDITIONS:**

Standard Test Method Used:	ASTM D 6978-05
Deviation From Standard Test Method:	Used 1" Permeation Cell
Analytical Method:	UV/VIS Spectrometry
Testing Temperature:	35.0°C ± 2.0
Collection System:	Closed Loop
Specimen Area Exposed:	5.067 cm <sup>2</sup>
Selected Data Points:	25/test
Number of Specimens Tested:	3/test
Location Sampled From:	Cuff area

**DETECTION METHOD OF CHEMICAL PERMEATION; UV/VIS ABSORPTION SPECTROMETRY:**

Instrument: Perkin Elmer UV/VIS Spectrometer Lambda 25  
 UV/VIS Absorption Spectrometry was used to measure the absorbance of test chemicals, which permeated through the specimens into the collection medium. The collection medium was circulated in a closed loop at 11 ml/minute of flow rate through the testing period. Data collection was performed according to the programmed schedule by means of UV Winlab software from the Perkin Elmer Corporation. The list of the characteristic wavelengths is shown below.  
 Table 3. Characteristic Wavelengths used in UV/VIS Absorption Spectrometry

TESTING CHEMOTHERAPY DRUGS	WAVELENGTH (nm)
Carmustine (BCNU), 3.3 mg/ml (3,300 ppm)	229
Cisplatin, 1.0 mg/ml (1,000 ppm)	199
Cyclophosphamide (Cytosan), 20.0 mg/ml (20,000ppm)	200
Cytarabine, 100 mg/ml (100,000 ppm)	272
Dacarbazine (DTIC), 10.0 mg/ml (10,000 ppm)	320
Doxorubicin Hydrochloride, 2.0 mg/ml (2,000 ppm)	232
Etoposide (Toposar), 20.0 mg/ml (20,000 ppm)	205
Fluorouracil, 50.0 mg/ml (50,000 ppm)	269
Ifosfamide, 50.0 mg/ml (50,000 ppm)	200
Methotrexate, 25 mg/ml, (25,000 ppm)	303
Mitomycin C, 0.5 mg/ml (500 ppm)	217
Mitoxantrone, 2.0mg/ml (2,000ppm)	242
Paclitaxel (Taxol), 6.0 mg/ml (6,000 ppm)	231
Thiotepa, 10.0 mg/ml (10,000 ppm)	199
Vincristine Sulfate, 1.0 mg/ml (1,000 ppm)	220

**SAMPLE CHARACTERISTICS:**


Table 4. Thickness characteristics for the tested specimens Glove sample identified as Nitrile Powder Free Glove, Blue, Lot# ZP161212

Testing Chemotherapy Drugs	Thickness (mm)			Average (mm)	Weight/Unit Area (g/m <sup>2</sup> )
	Sample 1	Sample 2	Sample 3		
Carmustine (BCNU)	0.046	0.049	0.047	0.047	52.3
Cisplatin	0.045	0.050	0.046	0.047	
Cyclophosphamide (Cytosan)	0.047	0.043	0.048	0.046	
Cytarabine	0.047	0.044	0.048	0.046	
Dacarbazine (DTIC)	0.048	0.049	0.045	0.047	
Doxorubicin Hydrochloride	0.050	0.052	0.046	0.049	
Etoposide (Toposar)	0.048	0.052	0.046	0.049	
Fluorouracil	0.048	0.047	0.045	0.047	
Ifosfamide	0.044	0.049	0.048	0.047	
Methotrexate	0.045	0.047	0.054	0.049	
Mitomycin C	0.047	0.050	0.044	0.047	
Mitoxantrone	0.046	0.047	0.053	0.049	
Paclitaxel (Taxol)	0.048	0.042	0.044	0.045	
Thiotepa	0.047	0.049	0.043	0.047	
Vincristine Sulfate	0.045	0.044	0.048	0.046	

**RESULTS:**

Table 5. Permeation Test Results on: Glove sample identified as Nitrile Powder Free Glove, Blue, Lot# ZP161212.

TEST CHEMOTHERAPY DRUG AND CONCENTRATION	MINIMUM BREAKTHROUGH DETECTION TIME (Specimen 1/2/3) (Minutes)	STEADY STATE PERM. RATE (Specimen 1/2/3) (µg/cm <sup>2</sup> /minute)	OTHER OBSERVATIONS
Carmustine (BCNU) 3.3 mg/ml (3,300 ppm)	14.7 (15.1,14.7,16.8)	0.6 (0.5,0.5,0.7)	Moderate swelling and no degradation
Cisplatin 1.0 mg/ml (1,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Cyclophosphamide (Cytosan) 20 mg/ml (20,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Cytarabine, 100 mg/ml (100,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Dacarbazine (DTIC) 10.0 mg/ml (10,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Doxorubicin Hydrochloride 2.0 mg/ml (2,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Etoposide (Toposar) 20.0 mg/ml (20,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Fluorouracil 50.0 mg/ml (50,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Ifosfamide, 50.0 mg/ml (50,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Methotrexate 25 mg/ml (25,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Mitomycin C 0.5 mg/ml (500 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Mitoxantrone, 2.0mg/ml (2,000ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation
Paclitaxel (Taxol) 6.0 mg/ml (6,000 ppm)	No breakthrough up to 240 min.	N/A	Moderate swelling and no degradation
Thiotepa 10.0 mg/ml (10,000 ppm)	58.8 (110.0,58.8,67.0)	0.5 (0.3,0.5,0.6)	Slight swelling and no degradation
Vincristine Sulfate 1.0 mg/ml (1,000 ppm)	No breakthrough up to 240 min.	N/A	Slight swelling and no degradation



Tiffany L. Heller  
Assistant Manager  
Pharmaceutical Services  
AKRON RUBBER DEVELOPMENT LABORATORY, INC.



Ana C. Barbur, M.S.,  
Manager  
Chemical, Microbiological and Pharmaceutical Services



中国认可  
国际互认  
检测  
TESTING  
CNAS L0604

**Test Report**

No.: QDHL2004003544MD\_EN

Date: MAY.18,2020

Page: 1 of 6

Attention: To check the authenticity of testing (inspection) report & certificate, please connect via internet at the telephone: (86-755)8103071442, or email: CH.DOCUMENT@sgs.com

Client name : DONG TAI CITY HUAYI GLOVES CO.,LTD  
 Client address : NO.36 NANZHUANG ROAD, EAST NEW DISTRICT, 224200  
 DONGTAI CITY, JIANGSU, PEOPLE'S REPUBLIC OF CHINA  
 Sample Description : DISPOSABLE NITRILE GLOVES  
 Lot No. : NOT PROVIDED  
 Lot Size : NOT PROVIDED  
 Sample Quantity : 360PCS

As above test item and its relevant information regarding to the submission are provided and confirmed by the applicant. SGS is not liable to either the test item or its relevant information, in terms of the accuracy, suitability, reliability or/and integrity accordingly.

\*\*\*\*\*

Sample Receiving Date : APR.29,2020  
 Test Performing Date : APR.29,2020 TO MAY.18,2020

\*\*\*\*\*



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/term-and-conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/term-and-conditions/term-e-document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

QD 7361591  
 www.sgs.com.cn  
 sgs.china@sgs.com



中国认可  
国际互认  
检测  
TESTING  
CNAS L6604

**Test Report**

No.: QDHL2004003544MD\_EN

Date: MAY.18,2020

Page: 2 of 6

Test Requested

1. BS EN 455-1:2000 MEDICAL GLOVES FOR SINGLE USE – PART 1:  
REQUIREMENTS AND TESTING FOR FREEDOM FROM HOLES (CLAUSE 5.1)
2. BS EN 455-2:2015 MEDICAL GLOVES FOR SINGLE USE – PART 2:  
REQUIREMENTS AND TESTING FOR PHYSICAL PROPERTIES (CLAUSE 4.2,  
4.3, 5.2, 5.3)
3. BS EN 455-3:2015 MEDICAL GLOVES FOR SINGLE USE—PART 3:  
REQUIREMENTS AND TESTING FOR BIOLOGICAL EVALUATION (CLAUSE  
4.4)

Result

Pass  
Pass  
Pass

Remark: - Unless otherwise stated the results shown in this test report refer only to the sample(s) tested. This document cannot be used for publicity, without prior written approval of the SGS.

SGS-CSTC Standards  
Technical Services (Qingdao)  
Co., Ltd.

*Jessica Gao*



Jessica Gao  
Approved Signatory



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction clauses defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

QD

7361592

[www.sgs.com.cn](http://www.sgs.com.cn)

[sgs.china@sgs.com](mailto:sgs.china@sgs.com)

### Test Report

No.: QDHL2004003544MD\_EN

Date: MAY.18,2020

Page: 3 of 6

#### Test Conducted:

- BS EN 455-1:2000 Medical gloves for single use – Part 1: Requirements and testing for freedom from holes

Number of test sample	: 200 Pieces
Sample size	: M
Number of non-conforming gloves	: 1pc

Clause	Test Items	Result
5	Watertightness test for detection of holes	---
5.1	Referee testing	Pass (See note 1)

Note : 1 Sample quantity: 200pcs, AQL:1.5, Ac:7, Re:8, Found:1. See refer photo.  
The sample selecting amount for this clause is deviated to 200 pcs as assessed by SGS.

- BS EN 455-2:2015 Medical gloves for single use – Part 2: Requirements and testing for physical properties

Number of test sample	: 26 Pieces
Type	: examination/procedure gloves b)
Size	: Examination/procedure gloves: M

Clause	Test Items	Result
4	Dimensions	---
4.2	Length	Pass (See result 1)
4.3	Width	Pass (See result 1)
5	Strength	---
5.2	Force at break	Pass (See result 2)
5.3	Force at break after challenge testing	Pass (See result 2)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of client's instructions, if any. The company's sole responsibility is to its client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

QD

7361593

[www.sgs.com.cn](http://www.sgs.com.cn)

[sgs.china@sgs.com](mailto:sgs.china@sgs.com)





中国认可  
国际互认  
检测  
TESTING  
CNAS L0604

## Test Report

No.: QDHL2004003544MD\_EN

Date: MAY.18,2020

Page: 4 of 6

### Result 1: Dimensions

Size	M	
No.	Length (mm)	Width (mm)
1	240	96
2	241	96
3	246	97
4	246	96
5	245	97
6	240	97
7	244	96
8	241	97
9	241	96
10	245	96
11	245	96
12	247	96
13	243	95
Standard requirement	≥240	95±10
Median value	244	96

### Result 2: Strength

Size: M			
Force at break (N)			
Before aging		After aging	
No.	/	No.	/
1	16.5	1	16.3
2	17.1	2	16.0
3	15.0	3	16.2
4	16.9	4	14.7
5	16.2	5	17.7
6	15.1	6	16.5
7	17.0	7	16.7
8	16.7	8	17.1
9	12.1	9	12.2
10	15.6	10	14.8
11	13.4	11	12.2
12	17.2	12	14.6
13	13.0	13	15.7
Standard requirement	≥6.0	Standard requirement	≥6.0
Median value	16.5	Median value	16.0

Attention: To check the authenticity of testing/inspection report & certificate, please contact our salespersons at telephone: (86) 755 8160 7144, or visit: CH.Detection@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/for/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/for/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

QD 7361594  
www.sgs.com.cn  
e\_sgs.china@sgs.com

Member of the SGS Group (SGS SA)

Test Report

No.: QDHL2004003544MD\_EN

Date: MAY.18,2020

Page: 5 of 6

3. BS EN 455-3:2015 Medical gloves for single use – Part 3: Requirements and testing for biological evaluation

Number of test sample	: 5 Pieces
Sample size	: M
Finishes of gloves	: Powdered-free gloves other than surgeon's gloves

Clause	Test Items	Result
4.4	Powder-free gloves	Pass (See note 1)

Note : 1 Test according to EN ISO 21171:2006, the average mass of powder per glove is 0.18mg. (Requirement: ≤2mg per powder-free glove)

Remark: The declaration of conformity is only based on the actual value of laboratory activity, measurement uncertainty of the results not take into account.

Sample Photo:

Received Sample



SGS authenticate the photo on original report only

Attention: To check the authenticity of testing instructions report & certificate, please contact us at telephone: (86-755) 82307544, or email: CN.Hongkong@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

QD 7361595  
www.sgs.com.cn  
sgs.china@sgs.com

SGS



中国认可  
国际互认  
检测  
TESTING  
CNAS L0604


Test Report

No.: QDHL2004003544MD\_EN

Date: MAY.18,2020

Page: 6 of 6

Refer Photo (Watertightness test):



\*\*\*End of Report\*\*\*

Attention: To check the authenticity of testing inspection report & certificate, please contact us at telephone: (86-755)83197144; or email: CN\_Doc@sgs.com



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.sgs> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/terms-and-conditions/Terms-a-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's Instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

QD 7361596

www.sgs.com.cn  
e: sgs.china@sgs.com

Member of the SGS Group (SGS SA)

# MANUFACTURING PLANT



Established in 2013, new recent investment of \$7 Million into extra production capacity.



Current production of over 45 Million nitrile gloves a year.



7,452 square meters and consisting of 3 production lines.



Fully compliant with EU requirements and regulations.



Focusing on top of the range equipment, improving technical processes and passing on cost savings to customers whilst never diminishing on quality.



Our research and development department produces innovative designs and focuses on new cutting-edge technology.

# MANUFACTURING PLANT



# FDA Registration & Device Listing

[New Search](#)

[Back To Search Results](#)

<b>Device Classification Name</b>	<a href="#">Polymer Patient Examination Glove</a>
<b>510(k) Number</b>	K131823
<b>Device Name</b>	PROPRIETARY / DONGTAI/ OTHER CLIENTS PRIVATE LABELING
<b>Applicant</b>	DONG TAI CITY HUAYI GLOVES CO., LTD. RM 1606 BLDG, 1 JIAN XIANG YUAN NO. 209 BEI SI HUAN ZHONG RD HAI DI Beijing, CN 100083
<b>Applicant Contact</b>	Chu Xiaoan
<b>Correspondent</b>	DONG TAI CITY HUAYI GLOVES CO., LTD. RM 1606 BLDG, 1 JIAN XIANG YUAN NO. 209 BEI SI HUAN ZHONG RD HAI DI Beijing, CN 100083
<b>Correspondent Contact</b>	Chu Xiaoan
<b>Regulation Number</b>	<a href="#">880.6250</a>
<b>Classification Product Code</b>	<a href="#">LZA</a>
<b>Date Received</b>	06/20/2013
<b>Decision Date</b>	12/09/2013
<b>Decision</b>	Substantially Equivalent (SESE)
<b>Regulation Medical Specialty</b>	General Hospital
<b>510k Review Panel</b>	General Hospital
<b>Summary</b>	<a href="#">Summary</a>
<b>Type</b>	Traditional
<b>Reviewed By Third Party</b>	No
<b>Combination Product</b>	No

# FDA Registration & Device Listing

https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRL/ri.cfm

U.S. Department of Health & Human Services

Follow FDA | En Español

**FDA U.S. FOOD & DRUG ADMINISTRATION**

Home Food Drugs Medical Devices Radiation-Emitting Products Vaccines, Blood & Biologics Animal & Veterinary Cosmetics Tobacco Products

## Establishment Registration & Device Listing

FDA Home Medical Devices Databases

1 result found for **Establishment Registration** or **Business Trade Name** : *DONGTAI CITY HUAYI GLOVES CO.,*

New Search

Establishment Name	Registration Number	Current Registration Yr
<a href="#">DONGTAI CITY HUAYI GLOVES CO., LTD.</a> CHINA	3017252891	2021
<ul style="list-style-type: none"> <li>• <a href="#">Polymer Patient Examination Glove - Powder-Free Nitrile Patient Examination Gloves</a></li> <li>• <a href="#">Patient Examination Glove, Specialty - Nitrile Patient Examination Gloves; Nitrile Patient Examination Gloves, Powder Free Blue Nitrile Examination Gloves, Tested For Use With Chemotherapy Drugs; Powder Free Blue Nitrile Examination Gloves, Tested For Use With Chemotherapy Drugs</a></li> </ul>	Manufacturer	
		Manufacturer

Can't find what you're looking for? [Try a new search](#)

Page Last Updated: 06/14/2021  
 Note: If you need help accessing information in different file formats, see [Instructions for Downloading Viewers and Players](#)

U.S. Department of Health & Human Services

Follow FDA | En Español

**FDA U.S. FOOD & DRUG ADMINISTRATION**

Home Food Drugs Medical Devices Radiation-Emitting Products Vaccines, Blood & Biologics Animal & Veterinary Cosmetics Tobacco Products

## Establishment Registration & Device Listing

FDA Home Medical Devices Databases

New Search Back To Search Results

**Establishment:**  
 DONGTAI CITY HUAYI GLOVES CO., LTD.  
 NO. 36 Nanzhuang Road, High-Tech Industrial Development Zone Of Dongtai, CN 224200  
**Registration Number:** 3017252891  
**FEI Number\*:** 3017252891  
**Status:** Active  
**Date Of Registration Status:** 2021

**Owner/Operator:**  
 DONGTAI CITY HUAYI GLOVES CO., LTD.  
 NO. 36 Nanzhuang Road, High-Tech Industrial Development Zone Of Dongtai, CN 224200  
**Owner/Operator Number:** 10073578

**Official Correspondent:**  
 Yun Gao  
 NO. 36 Nanzhuang Road, High-Tech Industrial Development Zone Of Dongtai, CN 224200  
**Phone:** 086-515-89517999

**US Agent:**  
 Hui Hong  
 ABMED SERVICE INC  
 1312 17th Street Suite 692  
 Denver, CO US 80202  
**Phone:** 303 8000162 Ext  
**Email:** info@abmed.com.cn

\* Firm Establishment Identifier (FEI) should be used for identification of entities within the imports message set

Page Last Updated: 03/22/2021  
 Note: If you need help accessing information in different file formats, see [Instructions for Downloading Viewers and Players](#).  
 Language Assistance Available: Español | 繁體中文 | Tiếng Việt | 한국어 | Tagalog | Русский | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | 日本語 | فارسی | English

Accessibility | Contact FDA | Careers | FDA Basics | FOIA | No FEAR Act | Nondiscrimination | Website Policies / Privacy